

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2023/03/23
Person Assigned	Christof Schadt
Submitter Name	Christof Schadt christof.schadt@brainlab.com
Submission Date	2022/10/06

Correction Number	CP 2272
Log Summary:	Cleanup ROI Module
Name of Standard	PS3.3
Rationale for Correction:	The attribute description RT ROI Observations Sequence (3006,0080) refers to a “ROI Module” which does not exist. Adapt the wording accordingly.
Correction Wording:	

In PS3.3, section C.8.8.8 RT ROI Observations Module update the following text

C.8.8.8 RT ROI Observations Module

The RT ROI Observations Module specifies the identification and interpretation of an ROI specified in the Structure Set Module and ROI Contour Module.

Table C.8-44. RT ROI Observations Module Attributes

Attribute Name	Tag	Type	Attribute Description
RT ROI Observations Sequence	(3006,0080)	1	Sequence of observations related to ROIs defined in the ROI Module Structure Set ROI Sequence (3006,0020) . One or more Items shall be included in this Sequence.
>Observation Number	(3006,0082)	1	Identification number of the Observation. The value of Observation Number (3006,0082) shall be unique within the RT ROI Observations Sequence (3006,0080).