

## DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2025-01-19
Person Assigned	Björn Nolte
Submitter Name	Björn Nolte
Submission Date	2024-08-07

Correction Number	CP-2447
Log Summary:	Pregnancy Status correct mismatch of VR and defined terms
Name of Standard	DICOM PS3.3
Rationale for Correction:	Pregnancy Status (0010,21C0) has VR Unsigned Short (US) but has the defined terms 0001 ... 0004. The leading zeroes cannot be encoded in US. Change the defined terms to 0001H ... 0004H, the hex values corresponding to decimal integers 1...4.
Correction Wording:	

*Table C.2-4. – Replace 0001 ... 0004 with 0001H ... 0004H*

**Table C.2-4. Patient Medical Module Attributes**

...		
Pregnancy Status	(0010,21C0)	Describes pregnancy state of Patient.  Enumerated Values: 0001H not pregnant 0002H possibly pregnant 0003H definitely pregnant 0004H unknown
...		

*Table C.7-4a. – Replace 0001 ... 0004 with 1 ... 4*

**Table C.7-4a. Patient Study Module Attributes**

...		
Pregnancy Status	(0010,21C0)	Describes pregnancy state of Patient.

		Enumerated Values: 0001 <u>H</u> not pregnant 0002 <u>H</u> possibly pregnant 0003 <u>H</u> definitely pregnant 0004 <u>H</u> unknown
...		