

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2024/06/03
Person Assigned	Wim Corbijn
Submitter Name	Wim Corbijn
Submission Date	2023/11/08

Correction Number	CP-2381
Log Summary: Replace attributes by Referenced SOP Instance macro for Intravascular OCT Series Module	
Name of Standard	PS3.3
<p>Rationale For Correction:</p> <p>The Intravascular OCT Series Module is including the SOP Class UID and SOP Instance UID instead of the standard macro for this. For consistency replace this by including the related macro.</p>	
Correction Wording:	

Update PS3.3 C.8.27.1 Intravascular OCT Series Module.

Attribute Name	Tag	Type	Attribute Description
Modality	(0008,0060)	1	Type of device, process or method that originally acquired the data used to create the images in this Series. Enumerated Values: IVOCT See Section C.7.3.1.1.1 for further explanation.
Series Number	(0020,0011)	1	A number that identifies this Series.
Referenced Performed Procedure Step Sequence	(0008,1111)	1C	Uniquely identifies the Performed Procedure Step SOP Instance to which the Series is related. Only a single Item shall be included in this Sequence. Required if a Performed Procedure Step SOP Class was involved in the creation of this Series.
>Include Table 10-11 “SOP Instance Reference Macro Attributes”			
>Referenced SOP Class UID	(0008,1150)	4	Uniquely identifies the referenced SOP Class.
>Referenced SOP Instance UID	(0008,1155)	4	Uniquely identifies the referenced SOP Instance.

Attribute Name	Tag	Type	Attribute Description
Presentation Intent Type	(0008,0068)	1	<p>Identifies the intent of the images that are contained within this Series.</p> <p>Enumerated Values:</p> <p>FOR PRESENTATION</p> <p>FOR PROCESSING</p>