

# DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2023/11/13
Person Assigned	Christof Schadt
Submitter Name	Jim.percy@elekta.com
Submission Date	2023/05/30

Correction Number	CP-2309
Log Summary: Add channel obstruction to Treatment Termination reasons	
Name of Standard	PS3.16
Rationale for Correction:	In IHE-RO discussion of content for the Brachy Treatment Record, knowledge that an obstruction or stoppage of the channel during a delivery had occurred was determined to be valuable, and not covered by existing codes.
Correction Wording:	

Add to PS 3.16, Table CID 9561 Treatment Termination Reason

**Table CID 9561. Treatment Termination Reason**

Coding Scheme Designator	Code Value	Code Meaning	SNOMED-RT ID	UMLS Concept Unique ID
.....				
<b>DCM</b>	<b><u>XXXXXX</u></b>	<b><u>Channel Obstruction</u></b>		

Add to PS 3.16, Table D-1. DICOM Controlled Terminology Definitions

**Table D-1. DICOM Controlled Terminology Definitions (Coding Scheme Designator "DCM" Coding Scheme Version "01")**

Code Value	Code Meaning	Definition
...		
<b><u>XXXXXX</u></b>	<b><u>Channel Obstruction</u></b>	<b><u>The brachytherapy channel is obstructed and the treatment in this channel cannot proceed.</u></b>