

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2023/11/13
Person Assigned	Christof Schadt
Submitter Name	Jim.percy@elekta.com
Submission Date	2023/04/20

Correction Number	CP-2310
Log Summary: Wording improvement for C.8.8.5.4 and use of Study in RTSS references	
Name of Standard	PS3.3
Rationale for Correction:	New readers of the standard that were not part of the early development need more explanation guidance to what to do to get past the mistaken use of SOP Class in a Study Reference.
Correction Wording:	

Update in PS3.3, C.8.8.5 Structure Set Module

Table C.8-41. Structure Set Module Attributes

Attribute Name	Tag	Type	Attribute Description
...			
>RT Referenced Study Sequence	(300,0012)	3	Sequence of Studies containing Series to be referenced. One or more Items are permitted in this Sequence. See Section C.8.8.5.4 <u>for the recommended way to handle the SOP Class required by this macro.</u>
>>Include <u>Table 10-11 "SOP Instance Reference Macro Attributes"</u>			
...			

C.8.8.5.4 SOP Class UID in RT Referenced Study Sequence

Since RT Referenced Study Sequence (3006,0012) is Type 3, the Attribute may be omitted.

If RT Referenced Study Sequence (3006,0012) is present with an Item, ~~the~~ **a SOP Class UID is required by the called Macro even though a Study can contain multiple SOP Classes. Therefore, if the Study contains Instances of different SOP Classes, it is recommended that implementers use of** the Detached Study Management SOP Class (Retired) **of 1.2.840.10008.3.1.2.3.1 for may be used in** Referenced SOP Class UID (0008,1150).