

DICOM Change Proposal

STATUS	Letter Ballot
Date of Last Update	2025/09/07
Person Assigned	Harry Solomon
Submitter Name	Detlev Hohmeier mailto:detlev.hohmeier@zeiss.com
Submission Date	2025/01/12

Change Number	CP-2499
Log Summary: Adjust Ophthalmic Photography Image Module Attribute Description for Referenced Image Sequence	
Name of Standard	PS3.3
Rationale for Change: The Ophthalmic Photography Image Module Attribute Descriptions needs to be corrected because since cp1559_ft_GeneralReferenceModule, Referenced Image Sequence (0008,1140) is not in the General Image Module (Section C.7.6.1) but in the General Reference Module (Section C.12.4).	
Change Wording:	

Modify PS3.3 Section C.8.17.2.1.1 Referenced Image Sequence as indicated

5 C.8.17.2.1 Ophthalmic Photography Image Module Attribute Descriptions

C.8.17.2.1.1 Referenced Image Sequence

The Referenced Image Sequence (0008,1140) in the General **ImageReference** Module (**Section C.7.6.1 Section C.12.4**) shall not convey stereoscopic information, which instead shall be encoded using the Stereometric Relationship IOD.