

DICOM Change Proposal

STATUS	Letter Ballot
Date of Last Update	2025/09/07
Person Assigned	steven.nichols@gehealthcare.com
Submitter Name	steven.nichols@gehealthcare.com
Submission Date	2025/02/24

Change Number	CP-2507
Log Summary:	Add Key Object Selection term to CID7010
Name of Standard	PS3.16
Rationale for Change:	IHE is introducing a named recovery option to IOCM in CP-RAD-423, and requires a new term to support this functionality.
Change Wording:	

Modify PS3.16 Section B. DCMR Context Groups as indicated (changes to existing text are bold and underlined for additions and bold and struckthrough for removals):

CID 7010 Key Object Selection Document Title

Resources: HTML | FHIR JSON | FHIR XML | IHE SVS XML
Keyword: KeyObjectSelectionDocumentTitle
FHIR Keyword: dicom-cid-7010-KeyObjectSelectionDocumentTitle
Type: Extensible
Version: ~~20190915~~ yyymmdd
UID: 1.2.840.10008.6.1.490

Table CID 7010. Key Object Selection Document Title

Coding Scheme Designator	Code Value	Code Meaning
DCM	113000	Of Interest
DCM	113001	Rejected for Quality Reasons
...		
<u>DCM</u>	<u>newCODE1</u>	<u>Rejection Withdrawn</u>

Modify PS3.16 D DICOM Controlled Terminology Definitions as indicated (changes to existing text are bold and underlined for additions and bold and struckthrough for removals):

D DICOM Controlled Terminology Definitions (Normative)

This Annex specifies the meanings of codes defined in DICOM, either explicitly or by reference to another part of DICOM or an external reference document or standard.

The contents of this table are available in OWL, XRDF and CSV format at <ftp://medical.nema.org/medical/dicom/resources/ontology/dcm/dcm.owl> and in [Bioportal](#).

Table D-1. DICOM Controlled Terminology Definitions (Coding Scheme Designator "DCM" Coding Scheme Version "01")

25

Code Value	Code Meaning	Definition	Notes
...			
113001	Rejected for Quality Reasons	Rejected for Quality Reasons	
...			
113037	Rejected for Patient Safety Reasons	List of objects whose use is potentially harmful to the patient. E.g., an improperly labeled image could lead to dangerous surgical decisions.	
113038	Incorrect Modality Worklist Entry	List of objects that were acquired using an incorrect modality worklist entry, and that should not be used, since they may be incorrectly identified.	
113039	Data Retention Policy Expired	List of objects that have expired according to a defined data retention policy.	
...			
<u>newCODE1</u>	<u>Rejection Withdrawn</u>	<u>List of objects that were previously rejected but whose rejection has been withdrawn, potentially restoring their accessibility.</u>	