

DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2023/09/02
Person Assigned	Kevin O'Donnell
Submitter Name	Kevin O'Donnell
Submission Date	2023/03/03

Correction Number	CP-2286
Log Summary:	Post-Coordinated Echo Property Code Fix
Name of Standard	PS3.16
Rationale for Correction:	<p>CID 12304 contains codes that express measured properties without pre-coordinating details that are encoded in other rows of the post-coordinated measurement template.</p> <p>Peak Blood Velocity should be present as a measured property, but not pre-coordinated with the Systolic timepoint which is encoded in a separate concept.</p>
Correction Wording:	

Amend DICOM PS3.16 CID 12304 Echo Measured Properties

Table CID 12304. Echo Measured Properties

Coding Scheme Designator	Code Value	Code Meaning	SNOMED-RT ID	UMLS Concept Unique ID	Units
...					
LN	20355-4 11726-7	Peak Systolic Blood Velocity		C0551845 C0803170	(m/s, UCUM, "m/s")
...					