

DICOM Change Proposal

| | |
|---------------------|-----------------------------------------------------------------------------------------|
| STATUS | Letter Ballot |
| Date of Last Update | 2025/11/10 |
| Person Assigned | Wim Corbijn |
| Submitter Name | Jeroen Medema, jeroen.medema@philips.com |
| Submission Date | 2025/01/13 |

| | |
|-----------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Change Number | CP-2500 |
| Log Summary: | Correct inconsistent Final State Requirement |
| Name of Standard | PS3.4 |
| Rationale for Change: | <p>The note in PS3.4, Section CC.2.5.1.3.2 says that the Unified Procedure Step Performed Procedure Sequence has a Final State requirement of X, while Table CC.2.5-3 preceding this note shows a Final State requirement of P for this attribute.</p> <p>The note is wrong in its conclusion.</p> |
| Change Wording: | |

Applicable text from the Table and Section highlighted below

5 CC.2.5.1.1 UPS Final State Requirements

COMPLETED and CANCELED are Final States for a UPS instance. The Attributes and values of the UPS instance must meet certain requirements before it may be placed in either of the Final States.

Note

10

A UPS instance is in the SCHEDULED state when created. See Section CC.1.1 for rules governing state transitions.

Attributes shall be valued as indicated by the Final State Codes in the Final State Column of Table CC.2.5-3 before the Procedure Step State (0074,1000) may be set to COMPLETED or CANCELED (i.e., Final State).

15

Performing systems are encouraged to ensure that the values for all Attributes reasonably reflect what was done and the Final State of the UPS. This may include blanking Attributes that are permitted to be empty and for which no reasonable value can be determined. The UPS contents should make it clear whether the step was completed, what work was done, what results were produced and whether the results are usable. See Section GGG.3.1 "What Was Scheduled Vs. What Was Performed" in PS3.17 for a discussion of methods to convey things like partial completion.

Note

20

The SCU may choose not to distribute, or otherwise make available, some or all instances created during the procedure step and referenced in the Output Information Sequence (0040,4033).

Table CC.2.5-1. Final State Codes

| Final State Code | Meaning |
|------------------|--------------------------------------------------------------------------------------------------|
| R | The UPS State shall not be set to COMPLETED or CANCELED if this Attribute does not have a value. |

| Final State Code | Meaning |
|------------------|---------------------------------------------------------------------------------------------------------------------------|
| RC | The UPS State shall not be set to COMPLETED or CANCELED if the condition is met and this Attribute does not have a value. |
| P | The UPS State shall not be set to COMPLETED if this Attribute does not have a value, but may be set to CANCELED. |
| X | The UPS State shall not be set to CANCELED if this Attribute does not have a value, but may be set to COMPLETED. |
| O | The UPS State may be set to either COMPLETED or CANCELED if this Attribute does not have a value. |

...

25 CC.2.5.1.3 UPS Attribute Service Requirements

...

Table CC.2.5-3. UPS SOP Class N-CREATE/N-SET/N-GET/C-FIND Attributes

| Attribute Name | Tag | Usage N-CREATE (SCU/SCP) | Usage N-SET (SCU/SCP) | Final State | Usage N-GET (SCU/SCP) | Matching Key Type | Return Key Type | Remark/Matching Type |
|-----------------------------------------------------|-------------|-----------------------------------|-----------------------|-------------|-----------------------|-------------------|-----------------|----------------------|
| ... | | | | | | | | |
| Unified Procedure Step Performed Procedure Sequence | (0074,1216) | 2/2 Shall be created empty | 3/2 | P | 3/2 | - | - | See CC.2.5.1.3.2. |
| ... | | | | | | | | |

...

CC.2.5.1.3.2 Unified Procedure Step Performed Procedure Sequence

- 30 The Attributes of the Unified Procedure Step Performed Procedure Sequence (0074,1216) shall only be retrieved with Sequence Matching.

Note

- 35 Since this Attribute may be created empty and has a Final State requirement of X, a UPS in the SCHEDULED state may be canceled with two N-ACTION requests (IN PROGRESS then CANCELED) and no N-SET request.